



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/619,428	07/16/2003	Takasuke Nakanishi	KAN-001 (10029)	6930
62479	7590	10/18/2007		
HAHN & VOIGHT PLLC 1012 14TH STREET, NW SUITE 620 WASHINGTON, DC 20005			EXAMINER MENDOZA, MICHAEL G	
			ART UNIT 3734	PAPER NUMBER
			MAIL DATE 10/18/2007	DELIVERY MODE PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

## Office Action Summary

**Application No.**

10/619,428

**Applicant(s)**

NAKANISHI, TAKASUKE

**Examiner**

Michael G. Mendoza

**Art Unit**

3734

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 01 June 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-5, 7, and 8 is/are pending in the application.
- 4a) Of the above claim(s) 8 is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-5 and 7 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)                                | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                       | 5) <input type="checkbox"/> Notice of Informal Patent Application                       |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____  |

## **DETAILED ACTION**

### ***Response to Arguments***

1. Applicant's arguments with respect to claims 1-5 have been considered but are moot in view of the new ground(s) of rejection.

### ***Election/Restrictions***

2. Newly submitted claim 8 directed to an invention that is independent or distinct from the invention originally claimed for the following reasons: Claim 8 is a method claim.

Since applicant has received an action on the merits for the originally presented invention, this invention has been constructively elected by original presentation for prosecution on the merits. Accordingly, claim 8 is withdrawn from consideration as being directed to a non-elected invention. See 37 CFR 1.142(b) and MPEP § 821.03.

### ***Claim Rejections - 35 USC § 102***

3. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:


A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

4. Claims 1-3, and 5 are rejected under 35 U.S.C. 102(b) as being anticipated by Simpson et al. 5047040.
5. Simpson et al. teaches a medical handpiece for cutting a treatment site, comprising: a cutting tool having an elongate flexible shank and a burr; a generally tubular sheath; an interposed member interposed between the elongate tube portion of

Art Unit: 3734

the sheath and the shank of the cutting tool; a hand piece body; wherein the elongate tube portion of the sheath is malleable and deformable, and the shank of the cutting tool and the interposed member are deformable following malleable deformation of the elongate tube portion; and wherein the shank has a first section to be chucked in a handpiece body, and a second section having a diameter smaller than the diameter of the first section having sufficient flexibility to be elastically deformed more easily than the first section; wherein the interposed member is tubular and flexible; wherein the interposed member is made of a fluoro-resin (col. 9, lines 44-47).



***Claim Rejections - 35 USC § 103***

6. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

7. Claim 4 is rejected under 35 U.S.C. 103(a) as being unpatentable over Simpson et al. in view of Brown 6514258.

8. Simpson et al. does not expressly disclose that his cutting tool has a marker that matches a marker of the same color of the sheath. However, Brown discloses that it is well known in the art to use markings on surgical tools to help the clinician know which bits to connect and the corresponding length of the cutting tool. It would have been obvious to one having ordinary skill in the art at the time the invention was made to use matching color markings, as disclosed by Brown, on the surgical device disclosed by Simpson et al. in order to help the clinician connect the correct cutting tool with the correct sheath or handpiece.

9. Claim 7 is rejected under 35 U.S.C. 103(a) as being unpatentable over Simpson et al.

10. Simpson et al. discloses the claimed invention except for the wall thickness range of the elongate tube portion. It would have been obvious to one having ordinary skill in the art at the time the invention was made to use the claimed range, since it has been held that wherein the general conditions of a claim are disclosed in the prior art,

Art Unit: 3734

discovering the optimum or workable ranges involves only routine skill in the art. *In re Aller*, 105 USPQ 233.

### **Conclusion**

11. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael G. Mendoza whose telephone number is (571) 272-4698. The examiner can normally be reached on Mon.-Fri. 9:00 a.m. - 5:00 p.m..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Michael Hayes can be reached on (571) 272-4959. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3734

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MM

A handwritten signature in black ink, appearing to read "M J Hayes", written in a cursive style.

MICHAEL J. HAYES  
SUPERVISORY PATENT EXAMINER